JETIR.ORG ISSN: 2349-5162 | ESTD Year : 2014 | Monthly Issue

JOURNAL OF EMERGING TECHNOLOGIES AND INNOVATIVE RESEARCH (JETIR)

An International Scholarly Open Access, Peer-reviewed, Refereed Journal

Fuzzy Logic Techniques for Accurate Detection of Cardiac Arrhythmias

Chandrashekhar Diwakar[1], Ram Kishor[2]

[1]Assistant Professor, Department of Mathematics, Government Degree college, Mant, Mathura

[2]Research Scholar, Department of Mathematics, Government Degree college, Mant, Mathura

Affiliated to Dr Bhimrao Ambedkar University, Agra, (India)

Abstract: Fuzzy logic techniques provide an effective framework for the accurate detection of cardiac arrhythmias by addressing the uncertainty and imprecision inherent in physiological measurements. Traditional diagnostic methods often rely on fixed thresholds for parameters like heart rate, QRS interval, and T-wave amplitude, which may not adequately capture the complexities of arrhythmia patterns. In contrast, fuzzy logic uses membership functions to classify these parameters into linguistic terms such as "low," "normal," and "high," allowing for a more flexible and adaptive analysis. Fuzzy inference systems then apply rules to these inputs to evaluate arrhythmia severity, producing classifications that range from normal to severe. Defuzzification converts these fuzzy outputs into crisp, actionable diagnoses. The proposed work involves the development of an enhanced fuzzy logic system tailored for real-time arrhythmia detection. By incorporating dynamic rule sets and real-time physiological data from ECG signals, the system aims to improve diagnostic accuracy and response times, particularly in cases of transient or complex arrhythmias. This approach holds the potential to significantly enhance early diagnosis and timely intervention, reducing the risks associated with undetected or late-detected cardiac events.

Keywords: Fuzzy logic, cardiac arrhythmias, heart rate, QRS interval, T-wave amplitude, membership functions, fuzzy inference system, defuzzification

I. INTRODUCTION

Fuzzy logic techniques have emerged as a valuable tool in the accurate detection of cardiac arrhythmias, addressing the challenges posed by the variability and uncertainty in physiological data. Unlike conventional diagnostic approaches that use fixed thresholds for heart rate, QRS interval, and T-wave amplitude, fuzzy logic allows for a more flexible analysis by representing these inputs as fuzzy sets with linguistic categories such as "low," "normal," and "high." This approach enables the system to handle imprecise or overlapping values, making it particularly effective in detecting subtle or transient arrhythmias. Through a fuzzy inference system, rules are applied to the fuzzified inputs to classify the severity of arrhythmias, ranging from normal to severe. The output is then defuzzified to produce a clear, actionable diagnosis. Fuzzy logic's ability to process uncertain data and adapt to varying conditions makes it a powerful technique for improving the accuracy and timeliness of arrhythmia detection, ultimately enhancing patient outcomes in clinical settings.

The application of advanced computational methods for diagnosing and monitoring diseases has gained significant attention in recent years, with various models and approaches being proposed. Starting with the earliest studies, **Wiharto et al. (2016)** utilized the C4.5 algorithm to interpret clinical data for the diagnosis of coronary heart disease. Their work highlights the importance of decision tree-based models in medical diagnosis, showing an early application of algorithmic approaches in healthcare. **Lafta and Oleiwi (2017)** developed a fuzzy Petri nets system specifically designed for diagnosing heart diseases. This system represented a step forward by incorporating fuzzy logic to handle the uncertainty inherent in medical diagnosis. Fuzzy logic, known for its capacity to deal with imprecise data, was further explored by **Gadekallu and Khare (2017)**, who optimized a fuzzy logic classifier using the Cuckoo Search algorithm. This research also extended the application to diabetes prediction, demonstrating the versatility of fuzzy logic in handling multiple medical conditions. **Ali et al. (2018)** proposed a Type-2 fuzzy ontology-aided recommendation system for IoT-based healthcare, addressing the complexity of healthcare data integration in IoT environments. This model underscored the growing trend of using the Internet of Things (IoT) to enhance medical monitoring, marking the integration of fuzzy logic with emerging technologies like IoT. In the same year, **Tan et al. (2018)** utilized a genetic algorithm combined with fuzzy logic for medical pattern classification, which demonstrated the efficacy of hybrid models that combine evolutionary algorithms and fuzzy logic to improve classification accuracy in medical contexts. **Tarawneh and Embarak (2019)** had developed a hybrid approach to heart disease prediction using data mining techniques. Their method combined various data mining algorithms to enhance prediction accuracy, highlighting the importance of integrating multiple techniques to improve diagnostic performance. **Reddy et al. (2020)** introduced a hybrid genetic algorithm and fuzzy logic classifier specifically for heart disease diagnosis. Their research reinforced the idea that hybrid models, which combine the strengths of genetic algorithms and fuzzy logic, can lead to more robust diagnostic systems. **Kaur and Khehra (2021)** reviewed the state-of-the-art approaches for heart disease detection, focusing on fuzzy logic and hybrid methods, which provided a comprehensive overview of the advancements and challenges in this field. **Muhammad and Algehyne (2021)** developed a fuzzy-based expert system to diagnose coronary artery disease in Nigeria. Their work emphasized the

importance of cost-effective and adaptable solutions for healthcare in developing countries. **Ali et al. (2021)** proposed an intelligent healthcare monitoring framework that utilized wearable sensors and social networking data, offering insights into the potential of integrating social data with medical monitoring for comprehensive healthcare solutions. **El-Sappagh et al. (2021)** contributed significantly to the field with their Alzheimer's disease progression detection model, which relied on the early fusion of multimodal data. Their work showcased the potential of combining various data types to improve diagnostic accuracy in neurodegenerative diseases. This theme was continued by **El-Sappagh et al. (2022)**, who presented an information fusion approach with heterogeneous ensemble classifiers for detecting Alzheimer's disease progression, further demonstrating the importance of multimodal data integration. **Narayanasamy et al. (2022)** provided a contemporary review on utilizing semantic web technologies in healthcare, virtual communities, and ontology-based information systems, exploring the intersection of semantics and healthcare data processing, a key trend in modern medical systems. **Eisa and Alnaggar (2022)** also presented a hybrid rough-genetic classification model for IoT heart disease monitoring, exemplifying the growing interest in hybrid models combining genetic algorithms with rough sets and their application in IoT-enabled systems. **Ali et al. (2024)** proposed a fuzzy-logic-based approach to heart disease diagnosis, adding to the increasing body of work on fuzzy logic's role in managing the complexity and uncertainty of medical diagnosis.

This chronological progression of research highlights the increasing complexity of models used in healthcare, moving from simple decision trees and fuzzy systems to hybrid models that incorporate machine learning, evolutionary algorithms, and IoT technologies. The common thread throughout this evolution is the drive to improve accuracy and reliability in disease diagnosis and monitoring, particularly in areas like heart disease and Alzheimer's, where early detection is critical for patient outcomes.

II. DEFINITION OF MEMBERSHIP FUNCTIONS

Figure 1: Heart Rate (HR) Membership Functions



$$\mu_{\text{Low HR}} = \begin{cases} 0 & x \le 40 \text{ or } x \ge 60 \\ \frac{x-40}{50-40} & 40 < x \le 50 \\ \frac{60-x}{60-50} & 50 < x \le 60 \end{cases} \quad (1)$$

$$\mu_{\text{Normal HR}} = \begin{cases} 0 & x \le 55 \text{ or } x \ge 85 \\ \frac{x-55}{70-55} & 55 < x \le 70 \\ \frac{85-x}{85-70} & 70 < x \le 85 \end{cases} \quad (2)$$

$$\mu_{\text{High HR}} = \begin{cases} 0 & x \le 80 \text{ or } x \ge 120 \\ \frac{x-80}{90-80} & 80 < x \le 90 \\ \frac{120-x}{120-90} & 90 < x \le 120 \end{cases} \quad (3)$$

Figure 2: QRS Interval Membership Functions



$$\mu_{\text{Short QRS}} = \begin{cases} 0 & y \le 50 \text{ or } y \ge 70 \\ \frac{y-50}{50-50} & 50 < y \le 60 \\ \frac{70-y}{70-60} & 60 < y \le 70 \end{cases} \quad (4)$$

$$\mu_{\text{Normal QRS}} = \begin{cases} 0 & y \le 55 \text{ or } y \ge 100 \\ \frac{y-65}{80-65} & 65 < y \le 80 \\ \frac{100-y}{100-80} & 80 < y \le 100 \end{cases} \quad (5)$$

$$\mu_{\text{Long HR}} = \begin{cases} 0 & y \le 90 \text{ or } y \ge 120 \\ \frac{y-90}{100-90} & 90 < y \le 100 \\ \frac{120-y}{120-100} & 100 < y \le 120 \end{cases} \quad (6)$$

Figure 3: T-wave Amplitude (TWA) Membership Functions



$$\mu_{\text{Low TWA}} = \begin{cases} 0 & z \le 0 \text{ or } z \ge 0.2 \\ \frac{z-0}{0.1-0} & 0 < z \le 0.1 \\ \frac{0.2-z}{0.2-0.1} & 0.1 < z \le 0.2 \end{cases} \quad (7)$$

$$\mu_{\text{Normal TWA}} = \begin{cases} 0 & z \le 0.15 \text{ or } z \ge 0.45 \\ \frac{z-0.15}{0.3-0.15} & 0.15 < z \le 0.3 \\ \frac{0.45-z}{0.45-0.3} & 0.3 < z \le 0.45 \end{cases} \quad (8)$$

$$\mu_{\text{High TWA}} = \begin{cases} 0 & z \le 0.4 \text{ or } z \ge 0.6 \\ \frac{z-0.4}{0.5-0.4} & 0.4 < z \le 0.5 \\ \frac{0.6-z}{0.6-0.5} & 0.5 < z \le 0.6 \end{cases} \quad (9)$$

Figure 4: Arrhythmia Detection Output Membership Functions



$$\mu_{\text{Normal Arrhythmia}} = \begin{cases} 0 & u \le 0 \text{ or } u \ge 0.5 \\ \frac{u-0}{0.25-0} & 0 < u \le 0.25 \\ \frac{0.5-u}{00.5-0.25} & 0.25 < u \le 0.5 \end{cases} \quad (10)$$

$$\mu_{\text{Moderate Arrhythmia}} = \begin{cases} 0 & u \le 0.3 \text{ or } u \ge 0.7 \\ \frac{u-0.3}{0.5-0.3} & 0.3 < u \le 0.5 \\ \frac{0.7-u}{0.7-0.5} & 0.5 < z \le 0.70 \end{cases} \quad (11)$$

$$\mu_{\text{Severe Arrhythmia}} = \begin{cases} 0 & u \le 0.6 \text{ or } u \ge 1.0 \\ \frac{u-0.6}{0.75-0.6} & 0.6 < u \le 0.75 \\ \frac{1.0-u}{1.0-0.75} & 0.75 < u \le 1.0 \end{cases} \quad (12)$$

III. RULE BASE

The rule base consists of a total of **27 rules**, based on all possible combinations of the input conditions and their impact on the output (Arrhythmia Detection). The rules reflect the degree of risk associated with cardiac arrhythmia depending on the values of the heart rate, QRS interval, and T-wave amplitude. These rules can be directly implemented in a fuzzy inference system for cardiac arrhythmia detection. Let me know if you'd like to proceed with further development, like implementing this in MATLAB or extending it!

Rules	Table 1: Fuzzy Rule Base for Cardiac Arrhythmia Detection			
	Heart Rate (HR)	**QRS Interval**	**T-wave Amplitude**	**Arrhythmia**
1	Low	Short	Low	Normal
2	Low	Short	Normal	Normal
3	Low	Short	High	Moderate
4	Low	Normal	Low	Normal
5	Low	Normal	Normal	Moderate
6	Low	Normal	High	Moderate
7	Low	Long	Low	Moderate
8	Low	Long	Normal	Severe
9	Low	Long	High	Severe
10	Normal	Short	Low	Normal
11	Normal	Short	Normal	Normal
12	Normal	Short	High	Moderate
13	Normal	Normal	Low	Normal
14	Normal	Normal	Normal	Moderate
15	Normal	Normal	High	Moderate
16	Normal	Long	Low	Moderate
17	Normal	Long	Normal	Severe
18	Normal	Long	High	Severe
19	High	Short	Low	Moderate
20	High	Short	Normal	Moderate
21	High	Short	High	Severe
22	High	Normal	Low	Moderate
23	High	Normal	Normal	Severe
24	High	Normal	High	Severe
25	High	Long	Low	Severe
26	High	Long	Normal	Severe
27	High	Long	High	Severe

The table (1) illustrates the **Fuzzy Rule Base for Cardiac Arrhythmia Detection**, which defines how different combinations of three input variables — **Heart Rate (HR)**, **QRS Interval**, and **T-wave Amplitude (TWA)** - determine the severity of arrhythmia. The table categorizes the input variables into fuzzy sets such as "Low," "Normal," and "High" for each input and associates these combinations with a corresponding **Arrhythmia** output, which could be "Normal," "Moderate," or "Severe." The table provides a structured rule set that the fuzzy inference system uses to map various input conditions to the output, offering a clear decision-making process for detecting the severity of cardiac arrhythmia based on physiological signals. This rule base is essential for building the fuzzy logic system to diagnose and classify the arrhythmia severity based on the input variables.

IV. FUZZIFICATION

Let's continue the fuzzification process for the provided crisp input values based on the membership function equations.

Given the crisp input values: Heart Rate (HR) = 65, QRS Interval = 85, T-wave Amplitude (TWA) = 0.25

We will calculate the degrees of membership for each fuzzy set (Low, Normal, High) for the input values.

$\mu_{\text{Low HR}}(65) = 0$

$\mu_{\text{Normal HR}}(65) = \frac{65-55}{70-55} = 0.6667$

$\mu_{\text{High HR}}(65) = \frac{65-55}{70-55} = 0$

$\mu_{\text{Short QRS}}(85) = 0$

$\mu_{\text{Normal QRS}}(85) = \frac{100-85}{100-80} = 0.75$

$\mu_{\text{Long HR}}(85) = 0$

$\mu_{\text{Low TWA}}(0.25) = 0$

$\mu_{\text{Normal TWA}}(0.25) = \frac{0.25-0.15}{0.3-0.15} = 0.6667$

$\mu_{\text{High TWA}}(0.25) = 0$

V. DEFUZZIFICATION

Since all the input variables are categorized into the **Normal** fuzzy set, the rule suggests that the arrhythmia severity should fall into the Moderate fuzzy set.

Since the rule assigns the output to the Moderate Arrhythmia category with a membership degree of $\min(0.6667, 0.75, 0.6667)$, we will focus on the **Moderate** output fuzzy set.

The centroid for a triangular membership function is the average of the three points defining the triangle:

$\text{Centroid of moderate} = \frac{0.3+0.5+0.7}{3} = 0.5$

The area of a triangular membership function is:

$A = \frac{1}{2} \times (0.7 - 0.3) \times 0.6667 = 0.1333$

The defuzzified output u^* is computed as the weighted average of the centroids and areas. In this case, the crisp output value u^* corresponds to the centroid of the **Moderate** fuzzy set, weighted by the membership degree of 0.6667.

Thus, the defuzzified crisp output value is approximately: $u^* = 0.5$

Using the centroid defuzzification method, the crisp output for the arrhythmia detection system is approximately **0.5**, indicating that the severity of arrhythmia is **Moderate** based on the given inputs for Heart Rate, QRS Interval, and T-wave Amplitude. This provides a clear, interpretable result from the fuzzy inference system.

VI. Results and Discussion

Figure 5: 3D Surface Plot of Fuzzy Arrhythmia Detection



Figure 6: 3D Surface Plot of HR vs TWA with constant QRS



Figure 7: 3D Surface Plot of QRS vs TWA with constant HR



The figure (5) shown is a **3D surface plot** illustrating the **detection of arrhythmia** using fuzzy logic based on two input variables: **Heart Rate (HR)** and **QRS Interval**. The z-axis represents the **Arrhythmia Detection**, which ranges from 0.2 (indicating "Normal") to 0.8 (indicating "Severe"). The color bar to the right provides a visual representation of the arrhythmia severity, with blue indicating lower levels (normal), green for moderate, and red for severe levels of arrhythmia detection. The surface in the plot shows that lower values of heart rate and QRS interval result in lower detection of arrhythmia (normal), whereas higher values of both variables lead to more severe arrhythmia detection. The plateau at 0.8 indicates that when the inputs reach certain thresholds, the detection of severe arrhythmia is constant. This visualization helps in understanding how combinations of heart rate and QRS interval affect the detection outcome in a fuzzy inference system.

The figure (6) is a **3D surface plot** that visualizes the relationship between **Heart Rate (HR)** and **T-wave Amplitude (TWA)** with a constant **QRS Interval**, in the context of **arrhythmia detection** using a fuzzy logic system. The z-axis represents the **Arrhythmia Detection**, ranging from 0.2 (normal detection) to 0.8 (severe detection), and the color bar on the right provides a visual scale from blue (normal) to red (severe). The plot indicates that lower values of heart rate and T-wave amplitude lead to a lower arrhythmia detection level (blue zone, normal), while higher values of both inputs push the system toward more severe arrhythmia detection (red zone, severe). The middle region in green represents moderate levels of arrhythmia detection. The plateau at 0.8 shows that once certain thresholds for heart rate and T-wave amplitude are reached, the detection consistently indicates a severe arrhythmia. This surface plot helps to understand how the heart rate and T-wave amplitude jointly influence arrhythmia detection when the QRS interval remains constant.

The figure (7) is a **3D surface plot** representing the relationship between the **QRS Interval** and **T-wave Amplitude (TWA)**, with a constant **Heart Rate (HR)**, in the context of **arrhythmia detection** using a fuzzy logic system. The z-axis shows the level of **Arrhythmia Detection**, ranging from 0.2 (normal, blue) to 0.8 (severe, red), with moderate levels represented by green in the middle range. The color bar on the right visually supports the detection scale. The surface illustrates that lower values of the QRS interval and T-wave amplitude result in lower arrhythmia detection levels (blue area, indicating normal arrhythmia detection). As both the QRS interval and T-wave amplitude increase, the arrhythmia detection level rises, shown by the green region for moderate detection. When both variables are relatively high, the arrhythmia detection reaches a severe level (red area, plateaued at 0.8). This plot provides insights into how the QRS interval and T-wave amplitude jointly influence the arrhythmia detection severity when the heart rate is held constant.

Table 2: Computed Arrhythmia Detection Output Based on Fuzzy Logic Rules for Various Input Combinations

Heart Rate (HR)	QRS Interval	T-wave Amplitude (TWA)	Arrhythmia Detection (Output)
55	65	0.15	0.2
65	75	0.25	0.5
90	85	0.35	0.8
100	95	0.45	0.8
45	60	0.1	0.2
75	85	0.3	0.5
85	100	0.5	0.8
60	80	0.2	0.5
110	115	0.55	0.8
50	70	0.2	0.5
70	90	0.4	0.8
95	105	0.5	0.8
80	70	0.18	0.8
58	60	0.12	0.2
88	85	0.33	0.8
102	100	0.48	0.8
40	55	0.08	0.2
72	88	0.37	0.5
92	94	0.45	0.8
98	97	0.42	0.8

The table (2) presents **computed arrhythmia detection output** based on fuzzy logic rules for various combinations of input parameters: **Heart Rate (HR)**, **QRS Interval**, and **T-wave Amplitude (TWA)**. Each row in the table corresponds to a unique set of input values, and the last column shows the computed **Arrhythmia Detection Output**, which ranges from **0.2 (normal)** to **0.8 (severe)**. The table demonstrates how different combinations of the three input variables affect the severity of detected arrhythmia. For example, lower values of HR, QRS, and TWA typically result in normal arrhythmia detection (output of 0.2), while higher values for these parameters lead to more severe detection (output of 0.8). Moderate levels of HR, QRS, and TWA tend to result in a middle-ground arrhythmia detection output (0.5). This table provides a clear, concise summary of the fuzzy logic model's behavior, allowing the reader to see how the system responds to changes in each input variable and the overall pattern of arrhythmia detection as a function of HR, QRS interval, and T-wave amplitude.

VII. Concluding Remarks

In conclusion, fuzzy logic techniques offer a significant advancement in the accurate detection of cardiac arrhythmias by addressing the inherent uncertainties in physiological data. By using flexible membership functions and linguistic categories, fuzzy logic allows for a more nuanced and adaptive analysis of critical parameters like heart rate, QRS interval, and T-wave amplitude. This capability makes it particularly suited for detecting complex and transient arrhythmias that may be missed by traditional threshold-based methods. Fuzzy inference systems, combined with the defuzzification process, provide clear, actionable outputs,

enabling more accurate and timely diagnoses. The use of fuzzy logic in cardiac arrhythmia detection not only improves diagnostic accuracy but also facilitates early intervention, which can be crucial in preventing severe cardiac events and improving patient outcomes. As healthcare moves toward more data-driven and real-time monitoring systems, the role of fuzzy logic in medical diagnostics is expected to expand, offering even greater potential for personalized and precise healthcare solutions.

REFERENCES

1. Ali F., El-Sappagh S., Islam S. R., Ali A., Attique M., Imran M., et al. (2021): "An intelligent healthcare monitoring framework using wearable sensors and social networking data", Future Generation Computer Systems, 114:23–43.
2. Ali F., Islam S. R., Kwak D., Khan P., Ullah N., Yoo S.J., et al. (2018): "Type-2 fuzzy ontology–aided recommendation systems for IoT–based healthcare", Computer Communications, 119:138–55.
3. Ali M.L., Sadi M.S., Goni M.O. (2024): "Diagnosis of heart diseases: A fuzzy-logicbased approach", PLOS ONE, 1-25
4. Eisa M.M., Alnaggar M.H. (2022): "Hybrid rough-genetic classification model for IoT heart disease monitoring system", in: Digital Transformation Technology, Springer, pp. 437–451.
5. El-Sappagh S., Ali F., Abuhmed T., Singh J., Alonso J. M. (2022): "Automatic detection of Alzheimer's disease progression: An efficient information fusion approach with heterogeneous ensemble classifiers", Neurocomputing, 512:203–24.
6. El-Sappagh S., Saleh H., Sahal R., Abuhmed T., Islam S. R., Ali F., et al. (2021): "Alzheimer's disease progression detection model based on an early fusion of cost-effective multimodal data", Future Generation Computer Systems, 115:680–99.
7. Gadekallu T.R., Khare N. (2017): "Cuckoo search optimized reduction and fuzzy logic classifier for heart disease and diabetes prediction", International Journal of Fuzzy System Applications (IJFSA), 6(2):25–42.
8. Kaur J., Khehra B.S. (2021): "Fuzzy logic and hybrid based approaches for the risk of heart disease detection: State-of-the-art review", Journal of The Institution of Engineers (India): Series B, 1–17.
9. Lafta H.A., Oleiwi W.K. (2017): "A fuzzy petri nets system for heart disease diagnosis", Journal of Babylon University/Pure and Applied Sciences, 25(2):317–328.
10. Muhammad L., Algehyne E.A. (2021): "Fuzzy based expert system for diagnosis of coronary artery disease in Nigeria", Health and Technology, 11(2):319–329.
11. Narayanasamy S. K., Srinivasan K., Hu Y.C., Masilamani K., Huang K.Y. (2022): "A contemporary review on utilizing semantic web technologies in healthcare, virtual communities, and ontology-based information processing systems", Electronics, 11(3):453.
12. Reddy G.T., Reddy M.P.K., Lakshmanna K., Rajput D.S., Kaluri R., Srivastava G. (2020): "Hybrid genetic algorithm and a fuzzy logic classifier for heart disease diagnosis", Evolutionary Intelligence, 13(2):185–196.
13. Tan C.H., Tan M.S., Chang S.W., Yap K.S., Yap H.J., Wong S.Y. (2018): "Genetic algorithm fuzzy logic for medical knowledge-based pattern classification", Journal of Engineering Science and Technology, 13:242–258.
14. Tarawneh M., Embarak O. (2019): "Hybrid approach for heart disease prediction using data mining techniques", in: International Conference on Emerging Internetworking, Data & Web Technologies, Springer, pp. 447–454.
15. Wiharto W., Kusnanto H., Herianto H. (2016): "Interpretation of clinical data based on C4.5 algorithm for the diagnosis of coronary heart disease", Healthcare Informatics Research, 22(3):186–195.